

Carbon Dioxide

Balance Gas	(M)SDS Range	(M)SDS	Regulator Recommendations	Page
Air	1 ppm - 50%	P-18-0012-A	Single Stage	PRS40023331-CGA E•245
Argon	1 ppm - 25%	P-18-0332	Two Stage	PRS40123331-CGA E•244
Helium	1 ppm - 50%	P-18-1641	Line	PRS40052001-000 E•246
Hydrogen	1 ppm - 20%	P-18-0102-A		
Nitrogen	0.01 ppm - 50%	P-18-0156-C		

Balance Gas	Concentration Range	CGA	Praxair Grades*						Cylinder Style	Pressure psig	Contents ft ³	m ³
			E	N	M	P	C	U				
Air	0.5 ppm - 50% ⁽¹⁾⁽²⁾	590	•	•	•	•	•	•	AS	2000	146	4.10
			•	•	•	•	•	•	AQ	2200	84	2.40
			•	•	•	•	•	•	A3	2200	32	0.90
			•	•	•	•	•	•	K	2000	199	5.64
			•	•	•	•	•	•	Q	2000	67	1.90
			•	•	•	•	•	•	G	2000	34	0.96
Argon	0.5 ppm - 50% ⁽¹⁾	580			•	•	•		AS	2000	146	4.10
					•	•	•		AQ	2200	84	2.40
					•	•	•		A3	2200	32	0.90
					•	•	•		K	2000	199	5.64
					•	•	•		Q	2000	67	1.90
					•	•	•		G	2000	34	0.96
Helium	0.5 ppm - 50% ⁽¹⁾	580			•	•	•		AS	2000	146	4.10
					•	•	•		AQ	2200	84	2.40
					•	•	•		A3	2200	32	0.90
					•	•	•		K	2000	199	5.64
					•	•	•		Q	2000	67	1.90
					•	•	•		G	2000	34	0.96
Hydrogen	0.5 ppm - 50% ⁽¹⁾	350			•	•	•		AS	2000	146	4.10
					•	•	•		AQ	2200	84	2.40
					•	•	•		A3	2200	32	0.90
					•	•	•		K	2000	199	5.64
					•	•	•		Q	2000	67	1.90
					•	•	•		G	2000	34	0.96
Nitrogen	0.5 ppm - 50% ⁽¹⁾⁽²⁾	580	•	•	•	•	•	•	AS	2000	146	4.10
			•	•	•	•	•	•	AQ	2200	84	2.40
			•	•	•	•	•	•	A3	2200	32	0.90
			•	•	•	•	•	•	K	2000	199	5.64
			•	•	•	•	•	•	Q	2000	67	1.90
			•	•	•	•	•	•	G	2000	34	0.96

⁽¹⁾ Cylinder pressure and contents will vary proportionately at higher concentrations.

⁽²⁾ Available concentrations for NIST traceable mixtures to be determined by available reference materials.

* **Key for Praxair Mixture Grades** (Refer to page C•92 for complete specification details.)

EPA (**E**), Primary Master (**PM**), Certified Master (**CM**), Primary Standard (**P**),
Certified Standard (**C**), Non-Certified Standard (**U**)