

Balance Gas	(M)SDS Range	(M)SDS	Regulator Recommendations	Page
Air	0.1 ppm - 0.5%	P-18-0067	Single Stage PRS40223331-CGA	E•248
Nitrogen	0.1 ppm - 0.5975%	P-18-0370-B	Two Stage PRS40323331-CGA	E•247
	0.5876 - 1%	P-18-0871-A	Line PRS40252001-CGA	E•249

Balance Gas	Concentration Range	CGA	Praxair Grades*						Cylinder Style	Pressure psig	Contents	
			E	N	M	P	C	U			ft ³	m ³
Air	2 ppm - 1% ⁽¹⁾	660	•	•	•	•	•	•	AS	2000	144	4.08
			•	•	•	•	•	•	AQ	2200	83	2.35
			•	•	•	•	•	•	A3	2200	31	0.88
Nitrogen	2 ppm - 1%	660				•	•	•	AS	2000	142	4.02
						•	•	•	AQ	2200	82	2.32
						•	•	•	A3	2200	31	0.88

⁽¹⁾ Available concentrations for NIST traceable mixtures to be determined by available reference materials.

* **Key for Praxair Mixture Grades** (Refer to page C•92 for complete specification details.)

EPA (**E**), Primary Master (**PM**), Certified Master (**CM**), Primary Standard (**P**),
Certified Standard (**C**), Non-Certified Standard (**U**)